

Addendum

Sub : Notice Inviting Tender for Supply, Delivery, Installation & Commissioning of 01 No. Multipara Cardic Monitor with EtCo2 Module for use in OT of Centenary Hospital of Syama Prasad Mookerjee Port Kolkata. Tender No. MED / PR / ADVT / 349 / 22-23/ 509

श्यामा प्रसाद मुखर्जी पोर्ट कोलकाता के शताब्दी अस्पताल के ओटी में उपयोग के लिए ईटीसीओ 2 मॉड्यूल के साथ 01 नंबर मल्टीपारा कार्डिक मॉनिटर की आपूर्ति, वितरण, स्थापना और चालू करने के लिए निविदा आमंत्रित करने की सूचना।

In terms of the Prebid meeting held on 20/5/2022 following has changes have been made in the NIT.

1. This Addendum shall be read in conjunction with the Bidding Document (including Notice Inviting Tender).
2. All other terms and conditions of the Bidding Document shall remain unchanged. Details of Corrigendum are given below :-


A: Following sl No of Technical Specification Appendix - B has been amended :


SL NO	Existing	Amended
1.2	Minimum 15 inches multi color TET display screen	Minimum 12 onward inches multi color TET display screen
1.3	Eight (Ten onwards) Channel digital and waveforms/ traces display. 10	Ten onwards Channel digital and waveforms/ traces display. 10
4.1	Multichannel (up to 12 lead) ST segment analysis.	Multichannel up to 3 / 5 lead ST segment analysis & QT & QTc analysis
5.1	Nellcor or Masimo technology.	Nellcor or Masimo technology. Global Technology Equivalent.
6.1	Should be Main Stream capnography with display of CO2 and digital Values of EtCO2, FiCO2 & RR.	Should be ARTEMA Main Stream -1Unit & Side Stream/Micro stream- 1 Unit and standard accessories in both model capnography with display of CO2 and digital Values of EtCO2, FiCO2 & RR.
15.1	Should be USFDA / CE approved product	Should be USFDA and European CE approved product
16		Should have option to CARDIAC OUTPUT upgradation module in future through thermolulution technology.(Added Point
17		Future upgradation to dual channel IBP Ports.
		NB : All accessories should be manufactured by the same company/ OEM

B. Tender Schedule: Tender schedule has been revised as follows:-

SL	Description	Existing	Amended
	Date of closing of Online e-tender for submission of Techno-Commercial Bid & Price Bid at Railtel Portal.	26/5/2022 15hrs	2/6/2022; 15hrs
	Date & time of opening of (Techno-Commercial Bid + Price bid) at RAILTELP Portal	26/5/2022	2/6/2022; 15:30 hrs

20/5/2022


 CHIEF MEDICAL OFFICER


 26/5/22